

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 5/11/2015

*Handwritten initials/signature*

Instrument Serial Number: 010519

As of 11-May-15

Start Date	End Date	Agency Name
06-May-15		DFS Central Lab
03-Dec-14	06-May-15	Franklin County SO
19-Nov-14	03-Dec-14	DFS Central Lab
03-Sep-14	19-Nov-14	Norfolk PD
19-Aug-14	03-Sep-14	DFS Central Lab
06-Mar-14	19-Aug-14	Halifax County Sheriff's Office
12-Apr-13	06-Mar-14	DFS Central Lab
26-Jul-12	12-Apr-13	Bedford County SO
20-Jun-12	26-Jul-12	DFS Central Lab
21-Jan-11	20-Jun-12	Bedford County SO
13-May-09	21-Jan-11	DFS Central Lab
04-Dec-08	13-May-09	Roanoke County Sheriff's Office
02-Jun-08	04-Dec-08	DFS Central Lab

# INTOX EC/IR II Quality Assurance Worksheet

*Handwritten initials/signature*

Instrument Serial Number 010519 Worksheet Start Date 5/6/2015  
 Location Franklin County  
 Address 40 E. Court St., Rocky Mount, VA 24151  
 DFS Technician Heather Stanton License No. 18910  
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 730 Reference Barometer (mm HG) 736  
 Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/16/2016

## Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.290
0.300	0.288	0.279	0.296	Sample 2	0.290
Precision		sample min	sample max	Sample 3	0.289
0.001		0.289	0.290		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.096
0.100	0.096	0.093	0.099	Sample 2	0.096
Precision		sample min	sample max	Sample 3	0.096
0		0.096	0.096		


Dry gas standard Lot No. (with tank no.) AG404902-15

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
 Mouthpieces  
 Certificates of Analysis  
 Operator Worksheet  
 Other:



Notes:

Instrument Serial Number

010519

Certification Date

11/11/15

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath M Stark

Date

5/11/15

Issuing Analyst

Jennie Duffy

Date

5/11/15




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

AMS  
JPD

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>STANTON, HEATHER, M</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>18910</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>05/06/2015</b>
TEST EQUIPMENT NUMBER <b>010519</b>		

RESULTS: TIME SAMPLE TAKEN 09:59 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH


ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 06 May 2015 at 10:26

HMS  
JPD

## Test Results

Instrument Serial Number 010519

Test # 000995 Subject Test

Test Location 1 Department of  
Test Date 06 May 2015

Test Location 2 Forensic Science  
Test Time 09:53  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name STANTON  
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 06 May 2015 End Time 10:00

Result Time 09:59

Result Date 06 May 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 09:53

Data Type BLK

Sample Value 0.000

Sample Time 09:54

Data Type CHK

Sample Value 0.096

Sample Time 09:55

Data Type BLK

Sample Value 0.000

Sample Time 09:56

Data Type SUBJ

Sample Value 0.000

Sample Time 09:56

Data Type BLK

Sample Value 0.000

Sample Time 09:57

Data Type SUBJ

Sample Value 0.000

Sample Time 09:59

Data Type BLK

Sample Value 0.000

Sample Time 10:00

Standard Type Dry Gas Std

Standard Value 0.096

Standard Lot Number AG404902-15

Standard Expiration Date 02/18/2016

Tank Pressure 806

Barometric Pressure 730 mmHg

Blow Sample Number 1 Blow Duration 2.97 sec

Blow Volume 1855 cc End-of-Blow Time 09:56

Blow Sample Number 2 Blow Duration 3.15 sec

Blow Volume 1828 cc End-of-Blow Time 09:59

Tamper Evident Stamp 0a81d1ea

Test Status Code 0

Test Status Success

HMS  
JPD

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010519      Test Number: 996

Test Date: 05/06/2015      Test Time: 10:01 EDT

Dry Gas Target: 0.096

Lot Number: AG404902-15      Exp Date: 02/18/2016

Tank Pressure: 801 psi      Barometric Pressure: 730 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:02
CHK	0.096	10:03
BLK	0.000	10:04
CHK	0.096	10:05
BLK	0.000	10:06
CHK	0.096	10:07

Test Status: *Success*

Calibration CRC: F12739EA

HMS  
JPD

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010519      Test Number: 997

Test Date: 05/06/2015      Test Time: 10:13 EDT

Dry Gas Target: 0.288

Lot Number: AG506501-08      Exp Date: 03/06/2017

Tank Pressure: 458 psi      Barometric Pressure: 730 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:14
CHK	0.290	10:15
BLK	0.000	10:16
CHK	0.290	10:17
BLK	0.000	10:19
CHK	0.289	10:20

Test Status: *Success*

Calibration CRC: F12739EA